LOGICA Medizintechnik GmbH	Technical Specification	FB4-01.4e Rev.01

No.:		Article category
55.37		Universal Adaptor with ENPlus-connector
Revision:	00	

Technical Property[1)]	Parameter[2)]
Connection for wide-necked flasks:	compatible with glass flasks with round thread acc. to DIN 168 - GL 40 x 4 compatible with ABBOTT® -Container
Connection for crown finish bottles:	compatible with crown finish 26 H 180 acc. to EN ISO 12821 compatible with crown finish 26 H 126 acc. to DIN EN 14635
Outlet with connecting thread for transfer sets:	cross connector acc. to ISO 18250
Opening pressure of air valve:	≤ 8 mbar
Hydrophobicity of air valve filter:	140 cm H2O
Air valve filter efficiency for 0.2 µm particle size:	approx. 99,99825 % (according to EN 1822-3)
Sterilization Compatibility[3)]:	ETO-Gas, ɣ-irradiation up to 50 kGy

Remarks:

LOGICA does not regulary provide biocompatibility testing data of its products. The distributor is solely responsible for procuring any required biocompatibility testing data.

1) LOGICA products are approved for single use. Multiple use may affect their technical properties adversely! LOGICA recommends validating the suitability for use of the selected parts in their specific application.
2) Specifications are for non-sterile products.
3) Sterilization of LOGICA products with sterilization methods which are not stated in this specification may affect their technical properties adversely!

	Prepared	Approved
Date:	20/01/2017	24.01.2017
Signature:		